

Foo et al.

S/N: 09/682,685

In the Claims

1. (Previously Presented) A method of multi-slice image acquisition with black-blood contrast comprising:

- applying a non-selective inversion pulse;
- applying a re-inversion pulse that is slice-selective over a region encompassing a plurality of slice selections;
- timing execution of a series of RF excitation pulses such that signal from blood is near a null point; and
- acquiring data for the plurality of slice selections.

2. (Original) The method of claim 1 wherein the plurality of slice selections include all slice selections in a slab to be imaged.

3. (Original) The method of claim 1 wherein the images are acquired over more than a single breath-hold.

4. (Original) The method of claim 1 wherein the re-inversion pulse is applied over a region having all slice selections in a slab and data are acquired for all slice selections in the slab using a single re-inversion pulse.

5. (Original) The method of claim 1 further comprising creating the inversion pulse with slice thickness given by:

$$\text{slice thickness} = (Z_1 - Z_n) + 4 * \text{opslthick},$$

where  $Z_1$  and  $Z_n$  represents spatial locations of first and last slices selected for imaging, and opslthick represents a desired imaging slice thickness.

6. (Original) The method of claim 5 further comprising creating the re-inversion pulse with a center centered about a midpoint between  $Z_1$  and  $Z_n$ .

7. (Original) The method of claim 1 wherein the timing step includes selecting an inversion time  $T_I$  such that the null point of the blood occurs near a center of the multi-slice acquisition.

Foo et al.

S/N: 09/682,685

8. (Currently Amended) The method of claim 1 ~~wherein the series of RF excitation pulses is fast spin echo readout pulses and wherein the method further comprises~~ modifying a flip angle of RF excitation pulses executed before and after an occurrence of the null point of the blood to improve blood suppression.

9. (Original) The method of claim 8 further comprising modifying the flip of RF excitation pulses occurring before the null point to slightly less than  $90^\circ$  and those occurring after the null point to slightly more than  $90^\circ$ .

10. (Currently Amended) A computer program stored on a computer readable storage medium and having a set of instructions that when executed by a computer cause the computer to:

(A) generate and cause application of a non-selective inversion RF pulse to a slab of slices each having a thickness;

(B) generate and cause application of a slice-selective re-inversion RF pulse having a slice thickness greater than the thickness of a single slice;

(C) apply an inversion time;

(D) apply RF excitations; and

(E) acquire MR data.

11. (Original) The computer program of claim 10 wherein the slice thickness of the re-inversion pulse is selected greater than the slab of slices to allow for cardiac motion between the application of the slice-selective re-inversion RF pulse, and the acquisition of MR data.

12. (Previously Presented) The computer program of claim 10 wherein the RF excitations have a flip angle greater than  $90^\circ$  for segments after a null point and less than  $90^\circ$  for segments before the null point.

13. (Currently Amended) The computer program of claim 10 wherein the ~~sequence acts (A) - (E) are carried out is applicable~~ over one or more R-R intervals.

Foo et al.

S/N: 09/682,685

14. (Original) The computer program of claim 10 wherein the MR data is acquired during mid-diastole of an R-R interval.

15. (Previously Presented) An MR apparatus to produce consistent contrast in image acquisition comprising:

a magnetic resonance imaging (MRI) system having a plurality of gradient coils positioned about a bore of a magnet to impress a polarizing magnetic field and an RF transceiver system and an RF switch controlled by a pulse module to transmit RF signals to an RF coil assembly to acquire MR images; and

a computer programmed to apply a pulse sequence having:

a non-selective inversion pulse to invert spins in a longitudinal direction across an entire slab of slices;

a slice-selective re-inversion pulse having an implied width at least as large as that of the non-selective inversion pulse; and

a series of excitation pulses spaced apart from the slice-selective re-inversion pulse by an inversion time.

16. (Original) The MR apparatus of claim 15 wherein the slice-selective re-inversion pulse of the pulse sequence is further defined as having a width greater than that of the non-selective inversion pulse to extend on either side of the non-selective inversion pulse.

17. (Original) The MR apparatus of claim 16 wherein the slice-selective re-inversion pulse extends approximately twice the nominal slice thickness on either side of the non-selective inversion pulse.

18. (Original) The MR apparatus of claim 15 wherein the inversion time of the pulse sequence is selected such that blood signal is close to a null point.

19. (Currently Amended) The MR apparatus of claim 18 wherein the series of excitation pulses ~~is of a fast spin echo readout type and have~~ therein excitation pulses with differing flip angles.

Foo et al.

S/N: 09/682,685

20. (Original) The MR apparatus of claim 19 wherein excitation pulses occurring near a mid-point of the series have a flip angle near  $90^\circ$  and excitation pulses occurring before a mid-point have a flip angle less than  $90^\circ$  and excitation pulses occurring after the mid-point have a flip angle more than  $90^\circ$ .

21. (Previously Presented) A pulse sequence for use in multi-slice MR data acquisition comprising:

a non-selective inversion pulse applicable to a slab of slices;

a slice-selective re-inversion pulse applicable to at least a number of slices in the slab of slices; and

a series of excitation pulses applicable to the at least a number of slices in the slab of slices after an inversion time.

22. (Original) The pulse sequence of claim 21 wherein the inversion time is selected to allow signal from blood in a mid-point of the at least a number of slices to approach a null point.

23. (Original) The pulse sequence of claim 21 wherein the at least a number of slices includes all slices in the slab of slices.

24. (Original) The pulse sequence of claim 21 wherein the at least a number of slices includes fewer slices than those in the slab of slices but more than one.

25. (Original) The pulse sequence of claim 21 wherein the at least a number of slices includes more slices than those in the slab of slices.

26. (Original) The pulse sequence of claim 21 wherein the non-selective inversion pulse has a thickness given by:

$$\text{slice thickness} = (Z_1 - Z_n) + 4 * \text{opslthick},$$

where  $Z_1$  and  $Z_n$  represents spatial locations of first and last slices selected for imaging, and opslthick represents a desired imaging slice thickness.

Foo et al.

S/N: 09/682,685

27. (Original) The pulse sequence of claim 26 wherein the slice-selective re-inversion pulse has a center centered about a mid-point between  $Z_1$  and  $Z_0$ .

28. (Currently Amended) The pulse sequence of claim 21 wherein the series of excitation pulses have varying flip angles and are ~~fast spin echo readout excitation pulses~~.

29. (Original) The pulse sequence of claim 28 wherein excitation pulses that occur before a mid-point of the series have a flip angle of less than  $90^\circ$ , those near the mid-point have a flip angle near or at  $90^\circ$ , and those that occur after the mid-point have a flip angle greater than  $90^\circ$ .